

CERTIFICATE OF ANALYSIS

PRODUCT: DNA-3000, DNA-3250

LOT NUMBER: 5817

EXPIRATION DATE: 01Jan2022

APPEARANCE: Clear Frozen Liquid


ASSAYS:

ELISA ABSORBANCES:

ANTIBODY	Antigen Concentration 0.01 µg/well	
	STANDARD	LOT 5817
Positive 1	0.531	0.538
Positive 2	1.474	1.496
Positive 3	2.062	2.928
Negative 1	0.078	0.072
Negative 2	0.093	0.104
Negative 3	0.001	0.002

CONCENTRATION: 1.0 mg/ml

ABSORBANCE at 260nm at 1:100: 0.209 A₂₆₀



Quality Control Director

1/22/2015
Date