

CERTIFICATE OF ANALYSIS

PRODUCT: DNA-3000, DNA-3250

LOT NUMBER: 6191

EXPIRATION DATE: 01Jun2025

APPEARANCE: Clear Frozen Liquid

ASSAYS:

ELISA ABSORBANCES:

ANTIBODY	Antigen Concentration 0.01 µg/well	
	STANDARD	LOT 6191
Positive 1	3.16	3.26
Positive 2	1.69	1.69
Positive 3	1.66	1.61
Negative 1	0.005	0.011
Negative 2	0.153	0.107
Negative 3	0.113	0.111

CONCENTRATION: 1.05 mg/ml

ABSORBANCE at 260nm at 1:100: 0.21 A₂₆₀



Quality Control Director

10/3/2018
Date