

## CERTIFICATE OF ANALYSIS

**Product Number:** HMP-0100

**Description:** Positive Control Sera: Human antibody against Myeloperoxidase

**Lot Number:** 4999

**Appearance:** Lyophilized

**Assays:**

**ELISA:**

Test	Results
SSA	2 EU/ml
SSB	3 EU/ml
SCL	2 EU/ml
Sm	2 EU/ml
Sm/RNP	1 EU/ml
Jo-1	1 EU/ml

*Interpretation of Results:*

<16 EU/ml: Negative

16-20 EU/ml: Equivocal

>20 EU/ml: Positive

N/A: Not Available

**ELISA:**

Test	Results
PR-3	4 EU/ml
MPO	96 EU/ml

*Interpretation of Results:*


<4 EU/ml: Negative

4.1-5.9 EU/ml: Equivocal

>6 EU/ml: Positive

**Infectious Disease Testing:**

This product has tested non-reactive for HBsAg, HIVAg,  $\alpha$ -HIV,  $\alpha$ -HBC, and  $\alpha$ -HCV.



Quality Director

11/11/2009  
Date