

CERTIFICATE OF ANALYSIS

Product Number: HMS-0100

Description: Positive Control Sera: Human antibody against Mitochondrial Antigen

Lot Number: 6299

Appearance: Lyophilized

Assays:

ELISA:

| Test | Results |
|--------|-----------|
| SSA | <20 EU/ml |
| SSB | <20 EU/ml |
| SCL | <20 EU/ml |
| Sm | <20 EU/ml |
| Sm/RNP | <20 EU/ml |
| Jo-1 | <20 EU/ml |

Interpretation of Results:

<20 EU/ml: Negative
20-30 EU/ml: Equivocal
>30 EU/ml: Positive
N/A: Not Available

IFA:

| Test | Results |
|------|---------|
| MITO | 3240 |

Interpretation of Results:

<40: Normal

Infectious Disease Testing:

This product has tested non-reactive for HBsAg, HIVAg, α -HIV, α -HBC, and α -HCV.

Saad Al Banaa
Quality Director

09/20/2022
Date